

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 10/01/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:21
INTERNAL STANDARD	VERIFIED	04:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:21
BLANK TEST	0.00	04:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:23
BLANK TEST	0.00	04:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:24
BLANK TEST	0.00	04:25

