

SOLUTION CHANGE

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 12/01/2022

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:08
INTERNAL STANDARD	VERIFIED	03:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:09
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:11
BLANK TEST	0.00	03:11

