

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 01/20/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:22
INTERNAL STANDARD	VERIFIED	01:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:22
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:24
BLANK TEST	0.00	01:25

