

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

CFO

Date: 01/26/2022

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:49
INTERNAL STANDARD	VERIFIED	10:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:49
BLANK TEST	0.00	10:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:50
BLANK TEST	0.00	10:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:52
BLANK TEST	0.00	10:53

