

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 02/09/2022

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35667  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:02
INTERNAL STANDARD	VERIFIED	08:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:03
BLANK TEST	0.00	08:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:04
BLANK TEST	0.00	08:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:05
BLANK TEST	0.00	08:06

