

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 03/09/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:05
INTERNAL STANDARD	VERIFIED	01:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:07
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:09

