

**SUPERVISOR TEST**

-----  
 ONEIDA CO.  
 SHERIFF'S DEPT  
 DATAMASTER dmt: 123906  
 -----

Date: 04/06/2022

OPERATOR NAME:  
 ROBERT S CORNISH  
 Permit #: 48068  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:15
INTERNAL STANDARD	VERIFIED	21:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:16
BLANK TEST	0.00	21:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:17
BLANK TEST	0.00	21:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:18
BLANK TEST	0.00	21:19

