

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/14/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:14
INTERNAL STANDARD	VERIFIED	02:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:14
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:15
BLANK TEST	0.00	02:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:17
BLANK TEST	0.00	02:18

