

SUPERVISOR TEST

 ONEIDA CO.
 SHERIFF'S DEPT
 DATAMASTER dmt: 123906

Date: 04/27/2022

OPERATOR NAME:
 NICHOLAS A MARCANO
 Permit #: 49335
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:56
INTERNAL STANDARD	VERIFIED	16:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:57
BLANK TEST	0.00	16:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:58
BLANK TEST	0.00	16:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:59
BLANK TEST	0.00	17:00

