

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 05/02/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:44
INTERNAL STANDARD	VERIFIED	21:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:46
BLANK TEST	0.00	21:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:47
BLANK TEST	0.00	21:48

