

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 05/04/2022

OPERATOR NAME:
NICHOLAS A MARCANO
Permit #: 49335
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:30
INTERNAL STANDARD	VERIFIED	10:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:30
BLANK TEST	0.00	10:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:31
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:32
BLANK TEST	0.00	10:33

CFO

