

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 05/11/2022

OPERATOR NAME:
NICHOLAS A MARCANO
Permit #: 49335
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:29
INTERNAL STANDARD	VERIFIED	17:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:29
BLANK TEST	0.00	17:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:31
BLANK TEST	0.00	17:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:32
BLANK TEST	0.00	17:33

CFO
~

