

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFU

Date: 05/18/2022

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:42
INTERNAL STANDARD	VERIFIED	23:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:43
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46

