

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 05/25/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:49
INTERNAL STANDARD	VERIFIED	19:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:50
BLANK TEST	0.00	19:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:51
BLANK TEST	0.00	19:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:52
BLANK TEST	0.00	19:53

