

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 06/15/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:28
INTERNAL STANDARD	VERIFIED	04:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:29
BLANK TEST	0.00	04:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:30

RFI Detected

