

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 07/13/2022

470

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:48
INTERNAL STANDARD	VERIFIED	00:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:49
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:50
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:51
BLANK TEST	0.00	00:52

