

SUPERVISOR TEST

 ONEIDA CO. CYD
 SHERIFF'S DEPT
 DATAMASTER dmt: 123906

Date: 07/20/2022

OPERATOR NAME:
 JASON L OSIKA
 Permit #: 35567
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:28
INTERNAL STANDARD	VERIFIED	02:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:28
BLANK TEST	0.00	02:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:30
BLANK TEST	0.00	02:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:31
BLANK TEST	0.00	02:32

