

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/04/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:21
INTERNAL STANDARD	VERIFIED	04:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:22
BLANK TEST	0.00	04:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:23
BLANK TEST	0.00	04:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:24
BLANK TEST	0.00	04:25

