

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/24/2022

OPERATOR NAME:
NICHOLAS A MARCANO
Permit #: 49335
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:20
INTERNAL STANDARD	VERIFIED	07:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24

CFO
~

π ⊕ ⊕ - ⊕ ⊕ ⊕ - ⊕ ? ⊕

