

SUPERVISOR TEST

 ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 09/21/2022

OPERATOR NAME:
 STEVEN P EILERS
 Permit #: 49321
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:10
INTERNAL STANDARD	VERIFIED	04:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:10
BLANK TEST	0.00	04:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:11
BLANK TEST	0.00	04:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:13
BLANK TEST	0.00	04:13

