

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 10/05/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:11
INTERNAL STANDARD	VERIFIED	20:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:14
BLANK TEST	0.00	20:14

