

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 10/12/2022

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:51
INTERNAL STANDARD	VERIFIED	05:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:51
BLANK TEST	0.00	05:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:52
BLANK TEST	0.00	05:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:54
BLANK TEST	0.00	05:55

