

**SUPERVISOR TEST**

-----  
 ONEIDA CO.  
 SHERIFF'S DEPT  
 DATAMASTER dmt: 123906  
 -----

Date: 10/20/2022

OPERATOR NAME:  
 BAILEY J DIXON  
 Permit #: 50814  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:56
BLANK TEST	0.00	02:57

