

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 11/02/2022

OPERATOR NAME:  
NOAH A DANQUER  
Permit #: 50812  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22260

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 21:50 |
| INTERNAL STANDARD       | VERIFIED | 21:50 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 21:51 |
| BLANK TEST              | 0.00     | 21:52 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 21:52 |
| BLANK TEST              | 0.00     | 21:53 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 21:53 |
| BLANK TEST              | 0.00     | 21:54 |

