

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/16/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:40
INTERNAL STANDARD	VERIFIED	00:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:43
BLANK TEST	0.00	00:44

