

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/23/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:51
INTERNAL STANDARD	VERIFIED	00:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:54
BLANK TEST	0.00	00:55

