

SUPERVISOR TEST

ONEIDA CO.
 SHERIFF'S DEPT
 DATAMASTER dmt: 123906

CPD

Date: 12/14/2022

OPERATOR NAME:
 STEVEN P EILERS
 Permit #: 49321
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:32
INTERNAL STANDARD	VERIFIED	02:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:33
BLANK TEST	0.00	02:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:34
BLANK TEST	0.00	02:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:35
BLANK TEST	0.00	02:36

