

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: Orieta County

Instrument Serial #: 123906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
2/16/22	21440	34.00	34.01	34.01	MP4189	.10	.10	.10	Eiles	NO
2/23/22	21440	34.01	34.00	34.01	MP4189	.10	.09	.10	Eiles	NO
3/1/22	21440	34.01	34.01	34.01	MP4189	.10	.10	.10	Osira	YES
3/9/22	21440	34.01	34.01	34.01	MP4189	.10	.10	.10	Eiles	NO
3/11/22	21440	34.01	34.01	34.01	MP4189	.10	.10	.10	Eiles	NO
3/24	21440	34.01	34.01	34.01	MP4189	.10	.10	.10	Cusack	NO
3/29	21440	34.01	34.01	34.01	MP4189	.10	.10	.10	Cornish	YES
4/1/22	22030	34.01	34.01	34.01	MP4189	.10	.10	.10	Cornish	YES
4/6	22030	34.01	34.01	34.01	MP4189	.10	.10	.10	Cornish	YES
4/14	22030	34.01	34.01	34.01	MP4189	.10	.10	.10	Eiles	NO
4/20/22	22030	34.00	34.01	34.00	MP4189	.10	.10	.10	Eiles	NO
4/27/22	22030	34.01	34.01	34.01	MP4189	.10	.10	.10	MARCANO	NO

* - Please enter a √ to denote simulator solution change taking place