

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: Orcida County
Sheriff's

Instrument Serial #:
173906

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
5/2/22	22030	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	N
5/4/22	22030	34.01	34.01	34.00	MP4189	.10	.10	.10	MARCANO	NO
5/11/22	22030	34.01	34.01	34.01	MP4189	.10	.10	.10	MARCANO	NO
5/18/22	22030	34.01	34.01	34.01	MP4189	.10	.10	.10	Cusack	NO
5/25/22	22030	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	No
6/01/22	22100	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	Yes
6/08/22	22100	34.01	34.01	34.01	MP4189	.10	.10	.10	OSIKA	NO
6/15/22	22100	34.01	34.01	34.01	MP4189	.10	.10	.10	Eilers	N
6/22/22	22100	34.01	34.01	34.01	MP4189	.10	.10	.10	OSIKA	N
6/29/22	22100	34.01	34.01	34.01	MP4189	.10	.10	.10	Cusack	N
7/1/22	22100	34.00	34.00	34.00	MP4189	.10	.10	.10	Sorrentino	Yes
7/6/22	22100	34.01	34.01	34.01	MP4189	.10	.10	.10	OSIKA	NO

* - Please enter a √ to denote simulator solution change taking place