

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 05/03/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

CFO
Weekly/Monthly

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:07 |
| INTERNAL STANDARD | VERIFIED | 01:07 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 01:08 |
| BLANK TEST | 0.00 | 01:09 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:09 |
| BLANK TEST | 0.00 | 01:10 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:10 |
| BLANK TEST | 0.00 | 01:11 |

