

SOLUTION CHANGE

ONEIDA CO.
 SHERIFF'S DEPT
 DATAMASTER dmt: 123906

Date: 06/14/2023

CFO

OPERATOR NAME:
 STEVEN P EILERS
 Permit #: 49321
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:01
INTERNAL STANDARD	VERIFIED	00:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:03
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:05

