

SOLUTION CHANGE

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 07/02/2023

CFO

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:13
INTERNAL STANDARD	VERIFIED	22:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:15
BLANK TEST	0.00	22:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:16
BLANK TEST	0.00	22:17

