

SOLUTION CHANGE

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 08/02/2023

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:35
INTERNAL STANDARD	VERIFIED	04:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:36
BLANK TEST	0.00	04:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:37
BLANK TEST	0.00	04:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:38
BLANK TEST	0.00	04:39

