

SOLUTION CHANGE

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

CFO

Date: 10/01/2023

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:15
INTERNAL STANDARD	VERIFIED	00:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19

