

SOLUTION CHANGE

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 11/01/2023

OPERATOR NAME:  
BAILEY J DIXON  
Permit #: 50814  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:41
INTERNAL STANDARD	VERIFIED	23:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:43
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45

