

SOLUTION CHANGE

 ONEIDA CO.
 SHERIFF'S DEPT
 DATAMASTER dmt: 123906

Date: 02/01/2023

OPERATOR NAME:
 BAILEY J DIXON
 Permit #: 50814
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:37
INTERNAL STANDARD	VERIFIED	02:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:37
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:38
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:39
BLANK TEST	0.00	02:40

