

SOLUTION CHANGE

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/01/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:21
INTERNAL STANDARD	VERIFIED	19:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:22
BLANK TEST	0.00	19:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:24
BLANK TEST	0.00	19:25

