

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/12/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:00
INTERNAL STANDARD	VERIFIED	02:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:01
BLANK TEST	0.00	02:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:03
BLANK TEST	0.00	02:04

