

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 05/10/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

CFE

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:32
INTERNAL STANDARD	VERIFIED	13:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:33
BLANK TEST	0.00	13:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:34
BLANK TEST	0.00	13:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:35
BLANK TEST	0.00	13:36

