

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 05/23/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:37
INTERNAL STANDARD	VERIFIED	22:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:37
BLANK TEST	0.00	22:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:39
BLANK TEST	0.00	22:40

