

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/08/2023

LFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:05
INTERNAL STANDARD	VERIFIED	00:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:08
BLANK TEST	0.00	00:09

