

SUPERVISOR TEST

CFD

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/21/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:30
INTERNAL STANDARD	VERIFIED	01:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:31
BLANK TEST	0.00	01:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:32
BLANK TEST	0.00	01:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:33
BLANK TEST	0.00	01:34

