

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 06/28/2023

OPERATOR NAME:  
NICHOLAS A MARCANO  
Permit #: 49335  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:01
INTERNAL STANDARD	VERIFIED	07:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:04
BLANK TEST	0.00	07:05

CFO

