

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 07/12/2023

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:14
INTERNAL STANDARD	VERIFIED	21:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:14
BLANK TEST	0.00	21:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:15
BLANK TEST	0.00	21:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:17
BLANK TEST	0.00	21:17

