

SUPERVISOR TEST

CFD

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 08/16/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:24
INTERNAL STANDARD	VERIFIED	00:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:25
BLANK TEST	0.00	00:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:26
BLANK TEST	0.00	00:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:27
BLANK TEST	0.00	00:28

