

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/30/2023

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:05
INTERNAL STANDARD	VERIFIED	05:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:05
BLANK TEST	0.00	05:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:07
BLANK TEST	0.00	05:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:08
BLANK TEST	0.00	05:09

