

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/06/2023

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:45
INTERNAL STANDARD	VERIFIED	21:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:46
BLANK TEST	0.00	21:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:47
BLANK TEST	0.00	21:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:48
BLANK TEST	0.00	21:49

