

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/14/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:56
INTERNAL STANDARD	VERIFIED	20:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:57
BLANK TEST	0.00	20:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:58
BLANK TEST	0.00	20:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:59
BLANK TEST	0.00	21:00

