

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/27/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:16
INTERNAL STANDARD	VERIFIED	22:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:17
BLANK TEST	0.00	22:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:18
BLANK TEST	0.00	22:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:19
BLANK TEST	0.00	22:20

