

SUPERVISOR TEST

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 10/04/2023

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:53
INTERNAL STANDARD	VERIFIED	23:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:53
BLANK TEST	0.00	23:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:54
BLANK TEST	0.00	23:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:55
BLANK TEST	0.00	23:56

